



KenGen

KENYA ELECTRICITY GENERATING COMPANY PLC

KGK-SID-04-2023

TENDER FOR SUPPLY, ASSEMBLY AND COMMISSIONING OF RESEARCH & DEVELOPMENT CENTER LABORATORY EQUIPMENT AND FURNITURE (CITIZEN CONTRACTORS)

RFx 5000012417

27th March, 2023

CLARIFICATION I:

In Accordance with the “Tender for Supply, Assembly and Commissioning of Research & Development Center Laboratory Equipment and Furniture.” KenGen hereby issues **Clarification I** as follows.

No	Tender ref.	Bidder’s Questions	KenGen Response
1	Page 78 Item No. 4: Magnetic Stirrer with Hot Plate	Please clarify whether you need one that is analog or the digital type.	It is the digital type
2	Page 79 Item No. 5: Ion Selective Electrode/Ammonium Electrode	From the specifications listed under this item, it is not clear whether you need a pH/ISE meter which should be supplied with an ammonium electrode. The specifications are giving pH & mV requirements with no mention of the ammonium electrode. Please clarify on what’s needed to enable bidders to offer an item that best suits your need.	pH/ISE meter with ammonium electrode pH/ISE meter should be supplied with an ammonium electrode.

BIDDER’S ACKNOWLEDGEMENT OF CLARIFICATION NO I.

We, the undersigned hereby certify that Clarification is an integral part of the document and has been incorporated in the Tender proposal.

Signed.....

Tenderer.....

Date.....