



KENYA ELECTRICITY GENERATING COMPANY PLC

KGN-GDD-063-2023

TENDER FOR SUPPLY OF HPHT DUAL CELL AND ATMOSPHERIC
CONSISTOMETERS FOR GDD

(Women Enterprises)

Dated: 2nd May, 2023

Clarification No.1

In accordance with the **Tender for Supply of HPHT Dual Cell and Atmospheric Consistometers for GDD** KenGen issues a **Clarification No. 1** as follows:

NO	Clarification Request	KENGEN'S RESPONSE
1.	Must the size of the HPHT Dual Cell consistometer be Dimensions: About (65" height x 30" width x 30" depth) or one is allowed to quote bigger/different dimensions?	Maintain within this dimensions because of the installation space in the cement lab.
2.	Must it be chandler or fann?	Yes.

SUPPLIER ACKNOWLEDGEMENT OF CLARIFICATION NO.1

We, the undersigned hereby certify that the Clarification No.1 is an integral part of the document and the alterations set out in Clarification has been incorporated in the Tender Proposal.

Signed.....

Tenderer.....

Date.....